

UT521 QA测试报告

使用仪器: 电阻箱

环境温湿度: 24.3°C/50%

功能	量程	输出值	准确度	最大允许误差	测试值	测试结果
接地电压	200Vac	1Vac	± (1%±4)	0.6~1.4Vac	1.1Vac	合格
		10Vac		9.5~10.5Vac	10.2Vac	合格
		30Vac		29.3~30.7Vac	30.1Vac	合格
		50Vac		49.0~51.0Vac	49.8Vac	合格
		100Vac		98.6~101.4Vac	100.4Vac	合格
		150Vac		148.1~151.9Vac	149.8Vac	合格
		175Vac		172.9~177.1Vac	175.6Vac	合格
		195Vac		192.6~197.4Vac	194.9Vac	合格
接地电阻	20Ω	1Ω	± (2%±0.1)	0.88~1.12Ω	1.02Ω	合格
		3Ω		2.84~3.16Ω	2.98Ω	合格
		5Ω		4.80~5.20Ω	5.02Ω	合格
		10Ω		9.70~10.30Ω	10.03Ω	合格
		15Ω		14.60~15.40Ω	14.99Ω	合格
		19Ω		18.52~19.48Ω	19.03Ω	合格
	200Ω	20Ω		19.5~20.5Ω	19.8Ω	合格
		50Ω		48.9~51.1Ω	50.2Ω	合格
		80Ω		78.3~81.7Ω	80.1Ω	合格
		100Ω		97.9~102.1Ω	100.2Ω	合格
		150Ω		146.9~153.1Ω	150.3Ω	合格
		190Ω		186.1~193.9Ω	189.8Ω	合格
	2KΩ	200Ω		193~207Ω	201Ω	合格
		350Ω		340~360Ω	348Ω	合格
		500Ω		487~513Ω	501Ω	合格
		750Ω		732~768Ω	748Ω	合格
		900Ω		879~921Ω	902Ω	合格
		1000Ω		977~1023Ω	1002Ω	合格
		1500Ω		1467~1533Ω	1498Ω	合格
		1700Ω		1663~1737Ω	1696Ω	合格
	1900Ω	1859~1941Ω		1904Ω	合格	



检测员: 张琼

审核员: 郭桂余