

4V4H/2HQA测试报告

仪器型号: LM530LD

检验项目及要求		首检			复检			备注	
外壳及手提带		OK			OK				
基座、微动手轮		OK			OK				
开关运转状况		OK			OK				
各控制键及指示灯		OK			OK				
调制状态		OK			OK				
H		左	中	右	左	中	右		
	直线度 <±1mm/7m	H1	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
		H2	\	\	\	\	\	\	
		H3	\	\	\	\	\	\	
		H4	\	\	\	\	\	\	
	精度:<±1mm/7m	±0.5mm			±0.5mm				
	水	H1-H2	\			\			
	平	H2-H3	\			\			
	对	H3-H4	\			\			
接	H4-H1	\			\				
激光线质		OK			OK				

检验项目及要求		首检			复检			备注
天顶:<±1mm/3m								
V		V1	V2		V1	V2		
	<±1mm/7m	上	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
		中	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
		下	0.3 mm	0.3 mm	0.3 mm	0.3 mm	0.3 mm	
	V1-V3, V2-V4: 90°		OK	OK		OK	OK	
	V1-V3, V2-V4: 180°		OK	OK		OK	OK	
下对点:<±1mm/1m		0.1mm			0.1mm			
报警状态		OK			OK			
其它		\			\			
结论:		合格			合格			

首检员: 刘爱

