

4V4H/2HQA测试报告

仪器型号: LM520LD

检验项目及要求		首检				复检				备注	
外壳及手提带		OK				OK					
基座、微动手轮		OK				OK					
开关运转状况		OK				OK					
各控制键及指示灯		OK				OK					
调制状态		OK				OK					
H		左	中	右		左	中	右			
	直线度 <±1mm/7m	H1	0.3 mm	0.3 mm	0.3 mm		0.3 mm	0.3 mm	0.3 mm		
		H2	\	\	\		\	\	\		
		H3	\	\	\		\	\	\		
		H4	\	\	\		\	\	\		
	精度:<±1mm/7m	±0.5mm				±0.5mm					
	水	H1-H2	\			\					
	平	H2-H3	\			\					
	对	H3-H4	\			\					
接	H4-H1	\			\						
激光线质		OK				OK					

检验项目及要求		首检				复检				备注	
天顶:<±1mm/3m											
V		V1	V2	V3	V4	V1	V2	V3	V4		
	<±1mm/7m	上	0.3 mm				0.3 mm				
		中	0.3 mm				0.3 mm				
		下	0.3 mm				0.3 mm				
	V1-V3, V2-V4: 90°		OK				OK				
	V1-V3, V2-V4: 180°		OK				OK				
下对点:<±1mm/1m		0.1mm				0.1mm					
报警状态		OK				OK					
其它		\				\					
结论:		合格				合格					

首检员: 刘爱

