

UT575B 测试报告										
使用设备: 5500 5080 5200			环境温湿度: 20.6℃/46 RH%							
功能	量程	测试值 (50HZ/60HZ)	准确度	最大允许误差	QA	读数范围			测试值	
									1#	
电阻	交流、直流、选择法档 0.01Ω-300.0kΩ	0.51Ω	±2%rdg±5dgt	0.06Ω	100%	0.45	~	0.57	0.51	
		5.01Ω		0.15Ω	100%	4.86	~	5.16	5.03	
		28.01Ω		0.61Ω	100%	27.4	~	28.62	28.1	
		50.0Ω	±2%rdg±3dgt	1.3Ω	100%	48.7	~	51.3	50.00	
		280.0Ω		5.9Ω	100%	274.1	~	285.9	281.00	
		500Ω		13Ω	100%	487	~	513	502	
		2800Ω		59Ω	100%	2741	~	2859	2798	
		5.00KΩ		±4%rdg±3dgt	0.23kΩ	100%	4.77	~	5.23	5.00
		20.00KΩ			0.83kΩ	100%	19.17	~	20.83	20.11
		100.0KΩ	4.3kΩ		100%	95.7	~	104.3	99.10	
		200.0KΩ	8.3kΩ		100%	191.7	~	208.3	204.20	
		双钳法 0.01Ω-100Ω	0.51Ω	±10%rdg±5dgt	0.1Ω	100%	0.41	~	0.61	0.49
	5.0Ω		1Ω		100%	4	~	5	4.80	
	10.0Ω		1.5Ω		100%	8.5	~	11.5	9.4	
	50Ω		10Ω		100%	40	~	60	50	
	100Ω		15Ω		100%	85	~	115	99	
	电流	0.00mA-600A	5.00mA	±2%rdg±3dgt	0.13mA	100%	4.87	~	5.13	4.98
50.0mA			1.3mA		100%	48.7	~	51.3	50.2	
500mA			13mA		100%	487	~	513	497	
5.00A			0.3A		100%	4.87	~	5.13	4.96	
50.0A			1.3A		100%	48.7	~	51.3	50.0	
500A			13A		100%	487	~	513	499	
电压	0.00V-100.0V	10.0V	±2%rdg±3dgt	0.5V	100%	9.5	~	10.5	9.9	
		50.0V		1.0V	100%	48.7	~	51.3	50.1	
		90.0V		2.1V	100%	87.9	~	92.1	90.3	

首检员: 刘爱

复检员: 郭海梅

